

-----  
Midmark M11 - v1.0.5  
Total Cycles: 807

Name: RITTER2  
Sterilizer ID: V2313228  
02/16/2022 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.4 F 0.0  
2:00 101.4 F 0.3  
4:00 119.0 F 2.3  
6:00 147.8 F 5.2  
8:00 175.5 F 9.3  
10:00 200.0 F 9.2  
12:00 216.5 F 7.5  
14:00 229.6 F 12.2  
16:00 231.8 F 8.8  
18:00 239.4 F 10.7  
20:00 244.9 F 13.3  
22:00 257.1 F 19.8  
24:00 268.1 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.9 F 29.8  
2:00 271.5 F 29.1  
3:00 272.7 F 29.9  
4:00 271.7 F 29.1

Min 271.4 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 198.6 F 0.0  
10:00 202.1 F 0.0  
15:00 212.1 F 0.0  
20:00 208.8 F 0.0  
25:00 211.5 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:17  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:03:25  
CYCLE COMPLETE

02/16/2022 11:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_