
Midmark M11 - v1.0.5
Total Cycles: 3909

Name: RITTER2
Sterilizer ID: V2313228
09/05/2024 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.6 F 0.0
2:00 98.7 F 0.0
4:00 101.3 F 0.2
6:00 133.9 F 2.1
8:00 164.9 F 4.9
10:00 184.8 F 7.9
12:00 213.9 F 11.7
14:00 219.1 F 5.8
16:00 227.3 F 10.0
18:00 232.9 F 9.4
20:00 237.9 F 11.5
22:00 236.7 F 9.4
24:00 248.7 F 15.3
26:00 259.2 F 21.3
28:00 268.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.8
1:00 272.7 F 30.5
2:00 271.5 F 29.7
3:00 272.8 F 30.3
4:00 271.4 F 29.9

Min 271.4 F 29.7
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 216.3 F 0.0
10:00 222.1 F 0.0
15:00 229.2 F 0.0
20:00 235.3 F 0.0
25:00 230.7 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 29:05
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:07:42
CYCLE COMPLETE

09/05/2024 03:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____