
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008185
8:43 02/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:56
STERILIZING 3:57
136.4 C 220kPa 3:57
136.0 C 220kPa 4:27
136.4 C 219kPa 4:57
135.9 C 220kPa 5:27
136.4 C 223kPa 5:57
136.4 C 223kPa 6:27
136.4 C 218kPa 6:57
136.4 C 222kPa 7:27
136.4 C 222kPa 7:27
VENTING 7:27
AIR DRYING 8:16
CYCLE COMPLETE 38:16

Digital Signature #
B5087B7093A55552
