

-----  
Midmark M11 - v1.0.5  
Total Cycles: 231

Name: RITTER2  
Sterilizer ID: V2313228  
07/21/2021 10:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
2:00 187.2 F 3.4  
4:00 198.1 F 8.3  
6:00 218.0 F 11.5  
8:00 228.9 F 9.1  
10:00 235.2 F 10.3  
12:00 242.4 F 12.9  
14:00 242.8 F 12.2  
16:00 255.7 F 18.8  
18:00 267.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.5  
2:00 271.7 F 29.4  
3:00 272.6 F 29.5  
4:00 271.5 F 29.1  
  
Min 271.5 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
5:00 220.0 F 0.0  
10:00 224.1 F 0.0  
15:00 219.3 F 0.0  
20:00 217.3 F 0.0  
25:00 218.7 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:40  
CYCLE COMPLETE

07/21/2021 12:04 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01