
Midmark M11 - v1.0.0
Total Cycles: 1159

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 06:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.0 F 0.0
4:00 84.2 F 0.0
6:00 114.7 F 0.1
8:00 144.5 F 0.1
10:00 172.0 F 0.3
12:00 194.4 F 2.3
14:00 206.7 F 5.1
16:00 214.6 F 8.1
18:00 225.0 F 11.4
20:00 227.0 F 6.1
22:00 236.2 F 10.2
24:00 238.0 F 9.3
26:00 244.3 F 12.6
28:00 242.2 F 11.2
30:00 254.3 F 17.4
32:00 264.9 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.5 F 28.7
2:00 271.6 F 28.1
3:00 272.5 F 28.7
4:00 271.6 F 28.0

Min 271.5 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 218.8 F 0.0
10:00 218.0 F 0.0
15:00 216.4 F 0.0
20:00 212.8 F 0.0
25:00 209.2 F 0.0
30:00 185.4 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:58
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:12:08
CYCLE COMPLETE

09/13/2021 08:04 AM
ACCEPTED: X
REJECTED:
Operator: U01