

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000081  
9:09 21/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:01  
STERILIZING 4:23  
136.0 C 221kPa 4:23  
136.4 C 224kPa 4:53  
136.4 C 223kPa 5:23  
136.4 C 223kPa 5:53  
136.0 C 221kPa 6:23  
136.3 C 223kPa 6:53  
136.4 C 224kPa 7:23  
136.4 C 224kPa 7:53  
136.4 C 224kPa 7:53  
VENTING 7:53  
AIR DRYING 8:46  
CYCLE COMPLETE 23:46  
-----

Digital Signature #  
1A882F4C531E0DD0  
-----