

-----  
Midmark M11 - v1.0.0  
Total Cycles: 890

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2021 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
2:00 215.2 F 9.6  
4:00 231.6 F 8.1  
6:00 236.2 F 8.6  
8:00 237.8 F 9.0  
10:00 254.8 F 17.7  
12:00 269.3 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.7 F 28.1  
2:00 271.9 F 28.0  
3:00 272.2 F 28.3  
4:00 272.7 F 28.6

Min 271.6 F 28.0  
Max 272.9 F 28.9

VENTING CHAMBER

C385: VENT MODE

\*\* ITEMS NOT STERILE \*\*

06/14/2021 08:56 AM