
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000757
12:48 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:23
STERILIZING 4:27
136.4 C 232kPa 4:27
136.3 C 232kPa 4:57
136.4 C 233kPa 5:27
136.4 C 233kPa 5:57
136.2 C 230kPa 6:27
136.3 C 233kPa 6:57
136.3 C 232kPa 7:27
136.3 C 231kPa 7:57
136.3 C 231kPa 7:57
VENTING 7:57
AIR DRYING 8:53
CYCLE COMPLETE 8:57

Digital Signature #
12806E7BDB02556D
