
Midmark M11 - v1.0.0
Total Cycles: 1320

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2021 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 162.8 F 2.5
4:00 181.1 F 6.2
6:00 200.8 F 10.6
8:00 217.7 F 4.8
10:00 232.5 F 9.8
12:00 237.0 F 8.7
14:00 243.2 F 12.7
16:00 242.0 F 11.3
18:00 255.0 F 18.1
20:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 29.1
2:00 271.7 F 28.6
3:00 272.6 F 29.1
4:00 271.6 F 28.4

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 224.3 F 0.0
10:00 219.2 F 0.0
15:00 207.3 F 0.0
20:00 212.9 F 0.0
25:00 214.8 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

10/29/2021 10:45 AM
ACCEPTED: X
REJECTED:
Operator: U02