
Midmark M11 - v1.0.5
Total Cycles: 2348

Name: RITTER2
Sterilizer ID: V2313228
05/10/2023 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.2 F 0.0
2:00 177.4 F 3.5
4:00 195.1 F 8.2
6:00 213.5 F 12.6
8:00 225.5 F 8.5
10:00 238.9 F 11.8
12:00 240.4 F 11.6
14:00 239.1 F 10.3
16:00 252.0 F 16.9
18:00 263.4 F 23.5
20:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.3
2:00 272.5 F 29.7
3:00 271.6 F 29.2
4:00 272.3 F 29.4

Min 271.5 F 29.0
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 216.4 F 0.0
10:00 239.2 F 0.0
15:00 237.6 F 0.0
20:00 226.8 F 0.0
25:00 224.5 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

05/10/2023 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____