
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000088
9:54 28/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:45
136.0 C 221kPa 4:45
136.2 C 223kPa 5:15
136.3 C 224kPa 5:45
136.4 C 225kPa 6:15
136.4 C 225kPa 6:45
136.4 C 224kPa 7:15
136.0 C 222kPa 7:45
136.4 C 225kPa 8:15
136.4 C 225kPa 8:15
VENTING 8:15
CYCLE ABORTED 9:04
STOP BUTTON PRESSED
NOT STERILE
END TIME 9:04

Digital Signature #
04F4D915BC9E8243
