
Midmark M11 - v1.0.0
Total Cycles: 1093

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2021 12:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.7 F 0.0
2:00 118.4 F 0.0
4:00 165.9 F 0.0
6:00 180.9 F 0.1
8:00 199.6 F 2.4
10:00 215.5 F 5.5
12:00 224.8 F 9.0
14:00 234.8 F 12.9
16:00 232.0 F 8.5
18:00 243.0 F 13.3
20:00 240.6 F 10.9
22:00 237.4 F 9.1
24:00 250.8 F 15.4
26:00 261.6 F 21.6
28:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 272.0 F 28.2
3:00 271.6 F 28.3
4:00 272.6 F 28.8

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 233.9 F 0.0
10:00 217.7 F 0.0
15:00 209.2 F 0.0
20:00 205.9 F 0.0
25:00 204.8 F 0.0
30:00 203.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 28:40
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:07:10
CYCLE COMPLETE

08/18/2021 01:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____