
Midmark M11 - v1.0.5
Total Cycles: 1018

Name: RITTER2
Sterilizer ID: V2313228
04/21/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.1 F 0.0
2:00 186.3 F 2.4
4:00 206.8 F 8.1
6:00 225.8 F 10.5
8:00 236.4 F 10.3
10:00 234.2 F 8.3
12:00 240.1 F 10.3
14:00 249.0 F 15.0
16:00 261.6 F 22.1
18:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.5 F 29.3
2:00 271.5 F 28.8
3:00 272.5 F 29.3
4:00 271.5 F 28.6

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 231.5 F 0.0
10:00 228.7 F 0.0
15:00 227.0 F 0.0
20:00 223.7 F 0.0
25:00 221.2 F 0.0
30:00 217.7 F 0.0

Drying Complete

FILLING: 2:13
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

04/21/2022 11:13 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____