

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1525

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2022 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 168.8 F 2.6  
4:00 176.8 F 6.5  
6:00 199.9 F 11.7  
8:00 219.3 F 6.8  
10:00 232.8 F 12.3  
12:00 235.1 F 9.9  
14:00 236.1 F 8.7  
16:00 249.1 F 15.5  
18:00 261.4 F 22.5  
20:00 272.4 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.8 F 29.9  
2:00 271.5 F 29.2  
3:00 271.8 F 30.0  
4:00 272.5 F 29.8  
  
Min 271.2 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 215.8 F 0.0  
10:00 223.0 F 0.0  
15:00 237.0 F 0.0  
20:00 225.2 F 0.0  
25:00 221.9 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

09/22/2022 01:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_