
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001982
14:41 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 2:55
136.4 C 227kPa 2:55
136.4 C 226kPa 3:25
136.4 C 227kPa 3:55
136.4 C 222kPa 4:25
136.1 C 223kPa 4:55
136.4 C 222kPa 5:25
136.4 C 224kPa 5:55
136.1 C 223kPa 6:25
136.1 C 223kPa 6:25
VENTING 6:25
AIR DRYING 7:20
CYCLE COMPLETE 7:38

Digital Signature #
CE18D547F52076D9
