
Midmark M11 - v1.0.0
Total Cycles: 2963

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2023 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.6 F 0.0
2:00 135.4 F 0.1
4:00 159.4 F 1.2
6:00 187.5 F 5.2
8:00 210.2 F 9.7
10:00 225.8 F 8.4
12:00 229.9 F 9.8
14:00 233.1 F 8.4
16:00 242.5 F 12.6
18:00 241.3 F 11.1
20:00 254.7 F 17.9
22:00 266.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.3 F 29.1
4:00 272.2 F 28.8

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 212.5 F 0.0
10:00 228.9 F 0.0
15:00 246.3 F 0.0
20:00 234.2 F 0.0
25:00 223.9 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:42
CYCLE COMPLETE

03/01/2023 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____