

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4494

Name: RITTER2  
Sterilizer ID: V2313228  
04/07/2025 04:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.6 F 0.0  
2:00 108.2 F 0.8  
4:00 128.8 F 3.0  
6:00 151.7 F 6.0  
8:00 175.0 F 9.5  
10:00 192.6 F 13.1  
12:00 207.0 F 4.9  
14:00 219.3 F 8.6  
16:00 229.6 F 12.6  
18:00 231.4 F 8.8  
20:00 241.3 F 12.4  
22:00 241.9 F 12.2  
24:00 253.0 F 17.9  
26:00 263.2 F 23.9  
28:00 272.8 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.3 F 30.2  
2:00 271.8 F 29.6  
3:00 271.0 F 29.6  
4:00 272.6 F 30.8

Min 271.0 F 29.3  
Max 273.5 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 194.7 F 0.0  
10:00 198.5 F 0.0  
15:00 196.1 F 0.0  
20:00 192.3 F 0.0  
25:00 189.0 F 0.0  
30:00 192.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 28:25  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:07:20  
CYCLE COMPLETE

04/07/2025 05:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_