
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006323
13:30 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:25
STERILIZING 3:17
136.0 C 220kPa 3:17
136.2 C 224kPa 3:47
136.4 C 225kPa 4:17
136.3 C 221kPa 4:47
136.2 C 223kPa 5:17
136.4 C 220kPa 5:47
136.1 C 221kPa 6:17
136.3 C 224kPa 6:47
136.3 C 224kPa 6:47
VENTING 6:47
AIR DRYING 7:36
CYCLE COMPLETE 10:26

Digital Signature #
306587120C1E45FC
