
Midmark M11 - v1.0.5
Total Cycles: 503

Name: RITTER2
Sterilizer ID: V2313228
10/29/2021 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 181.8 F 3.2
4:00 195.6 F 8.4
6:00 214.9 F 8.6
8:00 229.3 F 9.7
10:00 234.5 F 8.9
12:00 244.6 F 13.0
14:00 244.8 F 13.1
16:00 257.7 F 20.0
18:00 269.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.2 F 29.0
2:00 272.5 F 29.5
3:00 271.7 F 29.1
4:00 272.2 F 29.2

Min 271.5 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 233.4 F 0.0
10:00 219.3 F 0.0
15:00 226.4 F 0.0
20:00 219.7 F 0.0
25:00 221.5 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

10/29/2021 10:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____