
Midmark M11 - v1.0.0
Total Cycles: 1272

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.0 F 0.0
2:00 110.1 F 0.0
4:00 138.9 F 0.1
6:00 152.6 F 0.7
8:00 167.7 F 3.1
10:00 183.0 F 5.8
12:00 203.4 F 9.6
14:00 219.4 F 8.4
16:00 226.8 F 8.4
18:00 236.5 F 13.0
20:00 236.8 F 10.0
22:00 238.4 F 9.0
24:00 247.5 F 14.2
26:00 258.6 F 20.2
28:00 268.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 29.0
2:00 271.9 F 28.9
3:00 272.2 F 29.0
4:00 271.9 F 28.8

Min 271.5 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 197.2 F 0.0
10:00 202.8 F 0.0
15:00 208.0 F 0.0
20:00 215.3 F 0.0
25:00 221.8 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:08
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:07:28
CYCLE COMPLETE

10/18/2021 10:06 AM
ACCEPTED: X
REJECTED:
Operator: U01