

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2644

Name: RITTER2  
Sterilizer ID: V2313228  
07/13/2023 12:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
2:00 196.1 F 4.2  
4:00 210.0 F 9.9  
6:00 223.4 F 6.6  
8:00 237.4 F 12.1  
10:00 237.5 F 10.4  
12:00 237.4 F 9.8  
14:00 252.6 F 17.2  
16:00 264.9 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.7  
2:00 271.6 F 28.9  
3:00 272.5 F 29.9  
4:00 272.4 F 29.4  
  
Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
5:00 226.6 F 0.0  
10:00 228.4 F 0.0  
15:00 223.1 F 0.0  
20:00 220.7 F 0.0  
25:00 221.1 F 0.0  
30:00 222.0 F 0.0  
Drying Complete

FILLING: 2:30  
HEATING: 17:52  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:32  
CYCLE COMPLETE

07/13/2023 01:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_