
Midmark M11 - v1.0.5
Total Cycles: 930

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 185.4 F 0.6
4:00 200.8 F 5.5
6:00 218.8 F 11.1
8:00 230.2 F 7.8
10:00 241.8 F 12.5
12:00 243.4 F 12.4
14:00 243.0 F 11.9
16:00 256.5 F 19.0
18:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.6 F 29.4
2:00 272.8 F 29.9
3:00 272.0 F 29.1
4:00 272.9 F 29.8

Min 271.8 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 233.6 F 0.0
10:00 224.0 F 0.0
15:00 232.0 F 0.0
20:00 225.5 F 0.0
25:00 226.4 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

03/30/2022 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____