

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2663

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2022 03:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
2:00 186.6 F 3.6  
4:00 208.5 F 8.4  
6:00 221.2 F 8.3  
8:00 232.6 F 9.2  
10:00 240.4 F 9.4  
12:00 244.0 F 12.7  
14:00 242.9 F 11.7  
16:00 256.4 F 18.6  
18:00 268.2 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.0 F 28.9  
2:00 272.3 F 28.9  
3:00 271.5 F 28.2  
4:00 272.0 F 28.9  
  
Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 233.3 F 0.0  
10:00 241.2 F 0.0  
15:00 239.1 F 0.0  
20:00 228.9 F 0.0  
25:00 226.2 F 0.0  
30:00 227.0 F 0.0  
  
Drying Complete

FILLING: 1:46  
HEATING: 19:11  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:07  
CYCLE COMPLETE

12/08/2022 04:27 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_