
Midmark M11 - v1.0.5
Total Cycles: 3427

Name: RITTER2
Sterilizer ID: V2313228
03/01/2024 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.4 F 0.0
2:00 153.2 F 1.6
4:00 168.3 F 4.6
6:00 190.0 F 8.4
8:00 206.4 F 12.8
10:00 222.0 F 6.7
12:00 232.7 F 11.2
14:00 231.7 F 7.7
16:00 244.7 F 13.3
18:00 243.1 F 11.9
20:00 254.5 F 17.9
22:00 265.0 F 24.1
24:00 270.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.9 F 29.5
2:00 271.3 F 28.9
3:00 272.6 F 29.7
4:00 272.1 F 29.1

Min 271.2 F 28.9
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 236.9 F 0.0
10:00 218.3 F 0.0
15:00 187.5 F 0.0
20:00 183.4 F 0.0
25:00 180.3 F 0.0
30:00 178.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

03/01/2024 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____