
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004195
12:25 15/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:08
STERILIZING 3:08
136.0 C 220kPa 3:08
136.4 C 221kPa 3:38
136.4 C 224kPa 4:08
136.3 C 223kPa 4:38
136.2 C 222kPa 5:08
136.3 C 222kPa 5:38
136.4 C 225kPa 6:08
136.4 C 221kPa 6:38
136.4 C 221kPa 6:38
VENTING 6:38
AIR DRYING 7:28
CYCLE COMPLETE 9:30

Digital Signature #
1ABC3EF932DB3F5F
