
Midmark M11 - v1.0.5
Total Cycles: 737

Name: RITTER2
Sterilizer ID: V2313228
01/31/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.5 F 0.0
2:00 124.3 F 1.5
4:00 149.2 F 4.4
6:00 177.5 F 8.5
8:00 198.1 F 13.2
10:00 219.4 F 6.4
12:00 232.1 F 11.4
14:00 233.1 F 8.6
16:00 240.6 F 10.9
18:00 245.9 F 13.4
20:00 258.0 F 20.0
22:00 268.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 271.9 F 29.1
2:00 272.1 F 29.6
3:00 271.5 F 28.8
4:00 272.3 F 29.6

Min 271.2 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 227.5 F 0.0
10:00 194.9 F 0.0
15:00 191.1 F 0.0
20:00 187.5 F 0.0
25:00 218.3 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:04
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

01/31/2022 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____