

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3005

Name: RITTER 1  
Sterilizer ID: V2178198  
03/15/2023 07:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.2 F 0.0  
2:00 78.5 F 0.0  
4:00 89.1 F 0.0  
6:00 112.9 F 0.3  
8:00 148.0 F 1.9  
10:00 175.3 F 4.1  
12:00 198.5 F 7.3  
14:00 214.4 F 10.6  
16:00 222.0 F 7.1  
18:00 228.4 F 8.1  
20:00 237.4 F 12.3  
22:00 233.9 F 8.7  
24:00 242.1 F 10.9  
26:00 245.1 F 12.8  
28:00 256.8 F 19.1  
30:00 267.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.4 F 28.9  
2:00 271.5 F 28.5  
3:00 272.4 F 28.9  
4:00 271.4 F 28.3  
  
Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 213.1 F 0.0  
10:00 221.5 F 0.0  
15:00 230.0 F 0.0  
20:00 232.2 F 0.0  
25:00 228.9 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:23  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:21  
CYCLE COMPLETE

03/15/2023 08:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_