
Midmark M11 - v1.0.5
Total Cycles: 1078

Name: RITTER2
Sterilizer ID: V2313228
05/05/2022 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 167.0 F 3.2
4:00 192.8 F 8.1
6:00 211.1 F 11.4
8:00 227.1 F 9.0
10:00 235.0 F 10.2
12:00 242.1 F 12.6
14:00 242.1 F 11.7
16:00 255.4 F 18.6
18:00 267.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.5
2:00 271.4 F 28.8
3:00 272.7 F 29.6
4:00 271.9 F 28.9

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 209.9 F 0.0
10:00 220.7 F 0.0
15:00 222.3 F 0.0
20:00 216.9 F 0.0
25:00 213.7 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

05/05/2022 01:10 PM
ACCEPTED: X
REJECTED:
Operator: U02