
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
2.2uS / 1.4ppm
CYCLE NUMBER 000013
7:36 04/09/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:05
STERILIZING 3:06
136.4 C 220kPa 3:06
136.4 C 224kPa 3:37
136.3 C 224kPa 4:07
136.4 C 225kPa 4:37
136.1 C 223kPa 5:07
136.5 C 225kPa 5:37
136.4 C 223kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 6:37
VENTING 6:37
AIR DRYING 7:23
CYCLE COMPLETE 37:23

Digital Signature #
F9117263FDFE9953
