
Midmark M11 - v1.0.5
Total Cycles: 2447

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.1 F 0.0
2:00 201.1 F 3.9
4:00 218.5 F 9.5
6:00 224.1 F 7.5
8:00 235.6 F 11.6
10:00 235.8 F 9.7
12:00 236.8 F 8.9
14:00 250.2 F 15.9
16:00 262.9 F 23.3
18:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 29.0
3:00 272.6 F 30.0
4:00 272.3 F 29.4

Min 271.5 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 226.6 F 0.0
10:00 244.4 F 0.0
15:00 239.9 F 0.0
20:00 228.9 F 0.0
25:00 226.2 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

05/31/2023 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____