
Midmark M11 - v1.0.5
Total Cycles: 1198

Name: RITTER2
Sterilizer ID: V2313228
06/16/2022 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 171.7 F 2.9
4:00 186.4 F 7.5
6:00 207.5 F 12.9
8:00 223.4 F 8.1
10:00 236.9 F 13.3
12:00 238.3 F 11.2
14:00 237.9 F 10.0
16:00 251.9 F 16.9
18:00 263.4 F 23.9
20:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.8
2:00 271.6 F 29.0
3:00 272.3 F 29.8
4:00 272.2 F 29.6

Min 271.2 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 228.4 F 0.0
10:00 215.7 F 0.0
15:00 250.4 F 0.0
20:00 228.6 F 0.0
25:00 228.8 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

06/16/2022 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____