
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008315
9:05 13/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:09
STERILIZING 3:09
136.4 C 222kPa 3:09
136.0 C 221kPa 3:39
135.9 C 220kPa 4:09
136.4 C 217kPa 4:39
136.4 C 220kPa 5:09
136.3 C 220kPa 5:39
136.0 C 222kPa 6:09
136.4 C 222kPa 6:39
136.4 C 222kPa 6:39
VENTING 6:39
AIR DRYING 7:31
CYCLE COMPLETE 7:45

Digital Signature #
39C2D8ED3DD59078
