
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000101
10:47 26/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:17
STERILIZING 4:43
136.4 C 222kPa 4:43
136.4 C 224kPa 5:13
136.2 C 221kPa 5:43
136.2 C 223kPa 6:13
136.3 C 224kPa 6:43
136.4 C 225kPa 7:13
136.3 C 221kPa 7:43
136.3 C 223kPa 8:13
136.3 C 223kPa 8:13
VENTING 8:13
CYCLE ABORTED 8:57
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:57

Digital Signature #
A5CAE78ADCDC0F66
