
Midmark M11 - v1.0.0
Total Cycles: 1331

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2021 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 140.3 F 1.1
4:00 152.5 F 4.2
6:00 179.8 F 8.3
8:00 204.2 F 13.2
10:00 224.4 F 8.0
12:00 236.8 F 13.3
14:00 236.6 F 10.3
16:00 234.7 F 7.7
18:00 248.1 F 14.6
20:00 260.0 F 21.2
22:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 271.5 F 29.3
2:00 271.5 F 29.4
3:00 271.8 F 29.5
4:00 272.0 F 29.5

Min 271.2 F 29.0
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 230.8 F 0.0
10:00 230.7 F 0.0
15:00 228.4 F 0.0
20:00 231.4 F 0.0
25:00 228.5 F 0.0
30:00 231.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

11/02/2021 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U01