
Midmark M11 - v1.0.0
Total Cycles: 2874

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.4 F 0.0
2:00 187.1 F 3.5
4:00 194.8 F 8.1
6:00 215.3 F 13.3
8:00 226.2 F 8.4
10:00 237.9 F 11.6
12:00 240.2 F 11.3
14:00 238.3 F 9.7
16:00 252.6 F 16.7
18:00 264.3 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.2
2:00 272.1 F 28.6
3:00 271.6 F 28.6
4:00 272.5 F 29.3

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 223.8 F 0.0
10:00 233.7 F 0.0
15:00 228.3 F 0.0
20:00 225.9 F 0.0
25:00 224.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

02/08/2023 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____