
Midmark M11 - v1.0.0
Total Cycles: 3346

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 03:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 196.9 F 2.3
4:00 217.8 F 8.2
6:00 229.7 F 10.3
8:00 235.9 F 10.5
10:00 236.4 F 8.3
12:00 240.0 F 10.0
14:00 248.8 F 14.7
16:00 261.6 F 21.8
18:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.8
2:00 272.0 F 28.8
3:00 272.3 F 28.7
4:00 272.0 F 28.8

Min 271.6 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 225.7 F 0.0
10:00 226.6 F 0.0
15:00 228.7 F 0.0
20:00 231.6 F 0.0
25:00 227.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

06/06/2023 04:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____