
Midmark M11 - v1.0.0
Total Cycles: 4563

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2024 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.5 F 0.0
2:00 96.6 F 0.0
4:00 115.2 F 1.3
6:00 142.4 F 3.9
8:00 168.2 F 6.6
10:00 194.6 F 10.2
12:00 214.4 F 8.9
14:00 210.4 F 6.0
16:00 222.1 F 9.7
18:00 235.2 F 11.6
20:00 236.7 F 9.9
22:00 237.9 F 9.7
24:00 245.0 F 12.8
26:00 255.6 F 18.3
28:00 265.1 F 24.2
30:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 273.2 F 29.5
2:00 271.9 F 28.9
3:00 271.9 F 29.1
4:00 273.1 F 29.5

Min 271.4 F 28.7
Max 273.3 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 210.6 F 0.0
10:00 188.9 F 0.0
15:00 186.9 F 0.0
20:00 186.4 F 0.0
25:00 185.9 F 0.0
30:00 185.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

05/16/2024 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____