
Midmark M11 - v1.0.0
Total Cycles: 3814

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 05:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 187.9 F 2.8
4:00 195.0 F 6.6
6:00 217.4 F 11.2
8:00 225.4 F 5.4
10:00 233.3 F 10.0
12:00 239.2 F 8.9
14:00 242.6 F 12.3
16:00 241.5 F 10.9
18:00 253.8 F 17.1
20:00 264.4 F 23.4
22:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.6 F 28.1
3:00 272.3 F 29.0
4:00 272.1 F 28.5

Min 271.4 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 220.0 F 0.0
10:00 229.1 F 0.0
15:00 225.9 F 0.0
20:00 222.1 F 0.0
25:00 221.4 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 22:05
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

09/25/2023 06:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____