
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000022
7:11 09/01/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:28
STERILIZING 4:35
136.0 C 220kPa 4:35
136.2 C 221kPa 5:05
136.0 C 221kPa 5:35
136.4 C 223kPa 6:05
136.4 C 222kPa 6:35
136.4 C 223kPa 7:05
136.2 C 221kPa 7:35
136.2 C 222kPa 8:05
136.2 C 222kPa 8:05
VENTING 8:05
AIR DRYING 8:57
CYCLE COMPLETE 9:06

Digital Signature #
9D5B90C8DF57A920
