
Midmark M11 - v1.0.0
Total Cycles: 3701

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2023 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.9 F 0.0
2:00 179.8 F 3.4
4:00 184.1 F 7.3
6:00 207.0 F 11.7
8:00 220.8 F 6.1
10:00 232.0 F 11.0
12:00 231.9 F 7.5
14:00 244.1 F 13.4
16:00 243.0 F 12.0
18:00 254.7 F 18.1
20:00 265.2 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.1 F 29.2
2:00 272.2 F 29.0
3:00 271.1 F 28.5
4:00 272.4 F 29.4

Min 271.1 F 28.4
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 215.5 F 0.0
10:00 224.3 F 0.0
15:00 228.2 F 0.0
20:00 231.4 F 0.0
25:00 229.1 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:04
CYCLE COMPLETE

08/29/2023 04:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____