
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003584
9:53 03/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:41
136.5 C 220kPa 3:41
136.4 C 222kPa 4:11
136.4 C 222kPa 4:41
136.4 C 224kPa 5:11
136.4 C 224kPa 5:41
136.4 C 223kPa 6:11
136.2 C 221kPa 6:41
136.4 C 224kPa 7:11
136.4 C 224kPa 7:11
VENTING 7:11
AIR DRYING 8:00
CYCLE COMPLETE 8:08

Digital Signature #
828C7D512070D80A
