
Midmark M11 - v1.0.5
Total Cycles: 1435

Name: RITTER2
Sterilizer ID: V2313228
08/30/2022 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 177.3 F 0.1
4:00 187.3 F 4.2
6:00 211.0 F 9.2
8:00 219.4 F 9.2
10:00 236.2 F 11.2
12:00 234.5 F 9.1
14:00 238.4 F 9.7
16:00 248.3 F 15.0
18:00 260.5 F 21.9
20:00 271.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 29.9
2:00 272.5 F 29.6
3:00 271.7 F 29.1
4:00 271.2 F 29.2

Min 271.1 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 213.8 F 0.0
10:00 225.2 F 0.0
15:00 228.8 F 0.0
20:00 228.5 F 0.0
25:00 230.1 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

08/30/2022 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____