

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006638  
11:42 17/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 3:00  
STERILIZING 3:58  
136.4 C 223kPa 3:58  
136.4 C 224kPa 4:28  
136.3 C 221kPa 4:58  
136.4 C 224kPa 5:28  
136.2 C 222kPa 5:58  
136.3 C 223kPa 6:28  
136.4 C 223kPa 6:58  
136.4 C 223kPa 7:28  
136.4 C 223kPa 7:28  
VENTING 7:28  
AIR DRYING 8:18  
CYCLE COMPLETE 8:21  
-----

Digital Signature #  
8850EF0937EB960A  
-----