

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 000627  
13:42 02/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:44  
STERILIZING 3:58  
136.0 C 227kPa 3:58  
136.2 C 230kPa 4:28  
136.4 C 230kPa 4:58  
136.4 C 232kPa 5:28  
136.4 C 231kPa 5:58  
136.4 C 232kPa 6:28  
135.9 C 229kPa 6:58  
136.3 C 231kPa 7:28  
136.3 C 231kPa 7:28  
VENTING 7:29  
AIR DRYING 8:19  
CYCLE COMPLETE 23:19  
-----

Digital Signature #  
B83B9407EBB7C89A  
-----