
Midmark M11 - v1.0.5
Total Cycles: 4509

Name: RITTER2
Sterilizer ID: V2313228
04/14/2025 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 86.4 F 0.0
4:00 114.4 F 0.4
6:00 112.6 F 2.7
8:00 154.0 F 5.9
10:00 190.2 F 9.6
12:00 203.6 F 13.5
14:00 216.0 F 5.0
16:00 227.2 F 8.7
18:00 237.2 F 12.9
20:00 233.6 F 9.0
22:00 239.7 F 11.0
24:00 244.1 F 12.8
26:00 255.4 F 18.9
28:00 265.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.4 F 29.0
3:00 272.6 F 29.7
4:00 271.6 F 28.9

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 217.5 F 0.0
10:00 201.4 F 0.0
15:00 203.5 F 0.0
20:00 211.1 F 0.0
25:00 203.5 F 0.0
30:00 190.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 29:48
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

04/14/2025 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____