

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000787  
14:31 10/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 5:42  
STERILIZING 7:19  
136.2 C 221kPa 7:19  
135.9 C 219kPa 7:49  
136.3 C 222kPa 8:19  
136.4 C 223kPa 8:49  
136.2 C 222kPa 9:19  
136.4 C 222kPa 9:49  
136.0 C 220kPa 10:19  
136.3 C 222kPa 10:49  
136.3 C 222kPa 10:49  
VENTING 10:49  
AIR DRYING 11:58  
CYCLE COMPLETE 12:51  
-----

Digital Signature #  
4817C31D1108D7CD  
-----