
Midmark M11 - v1.0.0
Total Cycles: 4849

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2024 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 159.8 F 0.6
4:00 175.0 F 1.7
6:00 195.6 F 5.7
8:00 211.5 F 9.7
10:00 224.0 F 10.0
12:00 228.5 F 8.3
14:00 238.9 F 13.1
16:00 238.9 F 10.4
18:00 236.2 F 8.3
20:00 249.9 F 15.1
22:00 261.1 F 21.4
24:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 29.0
2:00 272.0 F 28.4
3:00 271.6 F 28.3
4:00 272.5 F 29.0

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 208.0 F 0.0
10:00 212.3 F 0.0
15:00 223.2 F 0.0
20:00 230.7 F 0.0
25:00 228.1 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

08/23/2024 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____