

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2274

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2022 01:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
2:00 198.3 F 1.1  
4:00 207.5 F 6.3  
6:00 222.6 F 11.9  
8:00 230.2 F 7.8  
10:00 241.6 F 13.4  
12:00 241.8 F 11.6  
14:00 239.8 F 10.4  
16:00 253.9 F 17.3  
18:00 265.7 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.6 F 29.0  
2:00 271.8 F 28.3  
3:00 271.9 F 28.7  
4:00 272.5 F 28.9  
  
Min 271.2 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 234.5 F 0.0  
10:00 227.3 F 0.0  
15:00 231.3 F 0.0  
20:00 229.6 F 0.0  
25:00 229.0 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

08/31/2022 02:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_