
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000776
7:03 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:08
STERILIZING 3:17
136.1 C 229kPa 3:17
136.4 C 233kPa 3:47
136.2 C 231kPa 4:17
136.4 C 230kPa 4:47
136.2 C 231kPa 5:17
136.4 C 232kPa 5:47
136.4 C 232kPa 6:17
136.2 C 232kPa 6:47
136.2 C 232kPa 6:47
VENTING 6:47
AIR DRYING 7:33
CYCLE COMPLETE 7:36

Digital Signature #
E5D3C572C3401F68
