
Midmark M11 - v1.0.5
Total Cycles: 4052

Name: RITTER2
Sterilizer ID: V2313228
10/24/2024 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.7 F 0.0
2:00 104.0 F 0.0
4:00 116.3 F 0.0
6:00 142.0 F 0.2
8:00 167.0 F 2.6
10:00 186.3 F 5.8
12:00 212.3 F 9.5
14:00 223.6 F 9.2
16:00 227.5 F 9.1
18:00 237.4 F 11.5
20:00 240.2 F 11.7
22:00 238.6 F 10.2
24:00 251.3 F 16.5
26:00 262.1 F 22.8
28:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 271.9 F 29.9
2:00 273.4 F 30.3
3:00 272.3 F 29.6
4:00 272.1 F 30.2

Min 270.9 F 28.9
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 187.2 F 0.0
10:00 207.9 F 0.0
15:00 216.6 F 0.0
20:00 221.1 F 0.0
25:00 228.4 F 0.0
30:00 234.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:33
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:07:05
CYCLE COMPLETE

10/24/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____